
Midmark M11 - v1.0.5
Total Cycles: 2210

Name: RITTER2
Sterilizer ID: V2313228
04/03/2023 03:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.1 F 0.0
2:00 183.9 F 4.2
4:00 208.2 F 9.7
6:00 221.5 F 6.5
8:00 235.5 F 12.0
10:00 237.2 F 10.4
12:00 237.5 F 9.8
14:00 252.6 F 17.3
16:00 265.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.2 F 29.5
2:00 271.5 F 29.3
3:00 272.3 F 29.6
4:00 271.4 F 28.9

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 237.9 F 0.0
10:00 226.9 F 0.0
15:00 226.9 F 0.0
20:00 229.6 F 0.0
25:00 223.5 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 17:40
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:10
CYCLE COMPLETE

04/03/2023 04:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____