
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001501
10:05 05/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:09
STERILIZING 3:17
136.4 C 220kPa 3:17
136.4 C 226kPa 3:47
136.3 C 225kPa 4:17
136.4 C 225kPa 4:47
136.3 C 224kPa 5:17
136.4 C 226kPa 5:47
136.4 C 226kPa 6:17
136.4 C 226kPa 6:47
136.4 C 226kPa 6:47
VENTING 6:47
AIR DRYING 7:37
CYCLE COMPLETE 11:30

Digital Signature #
D611C6FCD08A70C5
