
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002530
9:37 21/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:15
STERILIZING 3:21
136.4 C 224kPa 3:21
136.4 C 226kPa 3:51
136.4 C 225kPa 4:21
136.4 C 224kPa 4:51
136.3 C 224kPa 5:21
136.4 C 224kPa 5:51
136.2 C 223kPa 6:21
136.4 C 226kPa 6:51
136.4 C 226kPa 6:51
VENTING 6:51
AIR DRYING 7:43
CYCLE COMPLETE 14:11

Digital Signature #
C015699AC20C5369
