
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001314
11:05 21/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:02
STERILIZING 4:11
136.0 C 220kPa 4:11
136.2 C 223kPa 4:41
136.4 C 221kPa 5:11
136.2 C 223kPa 5:41
136.3 C 224kPa 6:11
136.3 C 224kPa 6:41
136.4 C 223kPa 7:11
136.4 C 226kPa 7:41
136.4 C 226kPa 7:41
VENTING 7:41
AIR DRYING 8:32
CYCLE COMPLETE 23:32

Digital Signature #
68A8C807C4DD4F53
