

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1374

Name: RITTER2  
Sterilizer ID: V2313228  
08/08/2022 10:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.3 F 0.0  
2:00 196.2 F 0.1  
4:00 211.5 F 3.3  
6:00 221.9 F 8.4  
8:00 232.7 F 13.5  
10:00 236.5 F 10.7  
12:00 234.8 F 8.8  
14:00 239.2 F 10.0  
16:00 248.9 F 15.1  
18:00 261.2 F 22.1  
20:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.4 F 29.7  
2:00 271.9 F 28.9  
3:00 272.5 F 29.7  
4:00 272.1 F 29.1

Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 219.1 F 0.0  
10:00 234.8 F 0.0  
15:00 226.5 F 0.0  
20:00 216.8 F 0.0  
25:00 211.4 F 0.0  
30:00 209.0 F 0.0  
Drying Complete

FILLING: 1:10  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

08/08/2022 11:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_