
Midmark M11 - v1.0.0
Total Cycles: 3219

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2023 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.0 F 0.0
2:00 196.9 F 1.1
4:00 209.2 F 6.3
6:00 224.3 F 11.6
8:00 229.6 F 7.4
10:00 240.5 F 12.9
12:00 239.3 F 10.6
14:00 237.0 F 9.0
16:00 251.4 F 15.9
18:00 262.9 F 22.6
20:00 272.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.8 F 28.3
2:00 272.5 F 28.7
3:00 272.1 F 28.8
4:00 271.6 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 228.3 F 0.0
10:00 242.8 F 0.0
15:00 228.1 F 0.0
20:00 219.9 F 0.0
25:00 214.9 F 0.0
30:00 210.6 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

05/11/2023 12:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____