
Midmark M11 - v1.0.5
Total Cycles: 1482

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.3 F 0.0
2:00 153.3 F 2.1
4:00 164.7 F 6.4
6:00 195.2 F 11.5
8:00 215.5 F 5.5
10:00 229.3 F 10.6
12:00 231.2 F 7.8
14:00 241.2 F 11.8
16:00 244.4 F 13.4
18:00 257.4 F 20.3
20:00 268.7 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.7 F 29.9
2:00 272.2 F 29.8
3:00 272.6 F 29.8
4:00 272.1 F 29.7

Min 271.8 F 29.3
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 215.3 F 0.0
10:00 211.6 F 0.0
15:00 218.3 F 0.0
20:00 213.5 F 0.0
25:00 210.7 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

09/13/2022 10:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____