

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001420  
13:16 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 5:15  
STERILIZING 6:38  
136.5 C 221kPa 6:38  
136.4 C 223kPa 7:09  
136.5 C 223kPa 7:39  
136.4 C 219kPa 8:09  
136.4 C 224kPa 8:39  
135.9 C 221kPa 9:09  
136.5 C 222kPa 9:39  
136.3 C 223kPa 10:08  
VENTING 10:08  
AIR DRYING 11:15  
CYCLE COMPLETE 11:43  
-----

Digital Signature #  
F1A445635FFB73B2  
-----