
Midmark M11 - v1.0.5
Total Cycles: 1039

Name: RITTER2
Sterilizer ID: V2313228
04/27/2022 07:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 77.5 F 0.0
4:00 80.8 F 0.0
6:00 91.3 F 0.0
8:00 150.4 F 0.3
10:00 175.0 F 2.1
12:00 197.0 F 5.3
14:00 211.4 F 8.8
16:00 223.0 F 12.6
18:00 227.7 F 7.5
20:00 237.5 F 12.0
22:00 236.1 F 9.3
24:00 240.5 F 10.4
26:00 246.6 F 13.7
28:00 257.8 F 19.9
30:00 268.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.4 F 29.3
2:00 271.2 F 28.8
3:00 272.3 F 29.3
4:00 271.3 F 28.7

Min 271.1 F 28.6
Max 272.4 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 203.7 F 0.0
10:00 209.3 F 0.0
15:00 202.7 F 0.0
20:00 206.1 F 0.0
25:00 217.6 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

04/27/2022 08:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____