
Midmark M11 - v1.0.0
Total Cycles: 4936

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2024 08:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 71.0 F 0.0
4:00 87.0 F 0.0
6:00 122.3 F 0.0
8:00 151.5 F 0.3
10:00 173.5 F 0.5
12:00 194.6 F 1.9
14:00 218.0 F 5.1
16:00 220.2 F 8.3
18:00 229.7 F 11.8
20:00 230.1 F 6.5
22:00 240.5 F 11.5
24:00 236.3 F 8.7
26:00 241.8 F 10.6
28:00 246.8 F 13.3
30:00 258.4 F 19.5
32:00 268.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.3 F 28.6
2:00 272.5 F 28.6
3:00 271.9 F 28.1
4:00 271.5 F 28.0

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 228.7 F 0.0
10:00 215.0 F 0.0
15:00 214.2 F 0.0
20:00 181.5 F 0.0
25:00 180.0 F 0.0
30:00 231.8 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 33:23
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:11:40
CYCLE COMPLETE

09/25/2024 09:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____