
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007357
10:59 10/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:21
STERILIZING 3:26
136.4 C 220kPa 3:26
136.4 C 225kPa 3:56
136.5 C 222kPa 4:26
136.3 C 224kPa 4:56
136.4 C 225kPa 5:26
136.4 C 224kPa 5:56
136.0 C 221kPa 6:26
136.4 C 225kPa 6:56
136.4 C 225kPa 6:56
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 37:46

Digital Signature #
6DC52DC905BDF774
