
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002547
11:35 28/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:58
STERILIZING 2:58
136.0 C 220kPa 2:58
136.4 C 226kPa 3:28
136.4 C 226kPa 3:58
136.3 C 224kPa 4:28
136.4 C 226kPa 4:58
136.4 C 223kPa 5:28
136.3 C 224kPa 5:58
136.4 C 227kPa 6:28
136.4 C 227kPa 6:28
VENTING 6:28
AIR DRYING 7:19
CYCLE COMPLETE 20:15

Digital Signature #
63DE01B236705E7B
