
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001162
8:31 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:30
STERILIZING 4:49
136.0 C 220kPa 4:49
136.3 C 220kPa 5:19
136.3 C 222kPa 5:49
136.2 C 223kPa 6:19
136.4 C 223kPa 6:49
136.2 C 220kPa 7:19
136.2 C 223kPa 7:49
136.3 C 223kPa 8:19
136.3 C 223kPa 8:19
VENTING 8:19
CYCLE ABORTED 8:49
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:49

Digital Signature #
37903976A60A41E6
