

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2581

Name: RITTER 1  
Sterilizer ID: V2178198  
11/17/2022 06:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.1 F 0.0  
2:00 187.5 F 4.3  
4:00 193.6 F 9.2  
6:00 218.7 F 7.4  
8:00 230.1 F 10.2  
10:00 231.5 F 7.2  
12:00 242.9 F 12.2  
14:00 244.3 F 12.7  
16:00 257.5 F 19.7  
18:00 269.3 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.5 F 29.3  
2:00 271.4 F 28.6  
3:00 272.4 F 29.5  
4:00 272.0 F 29.0  
  
Min 271.3 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.7 F 0.0  
5:00 222.3 F 0.0  
10:00 222.9 F 0.0  
15:00 231.0 F 0.0  
20:00 228.2 F 0.0  
25:00 227.9 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:01  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

11/17/2022 07:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_