
Midmark M11 - v1.0.0
Total Cycles: 2345

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2022 02:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 188.9 F 3.8
4:00 190.3 F 8.8
6:00 223.9 F 8.3
8:00 232.0 F 9.6
10:00 235.3 F 8.2
12:00 243.5 F 13.0
14:00 242.9 F 11.9
16:00 256.6 F 19.1
18:00 268.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.6 F 29.3
2:00 271.6 F 28.6
3:00 271.6 F 28.9
4:00 272.4 F 29.2

Min 271.1 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.7 F 0.0
5:00 236.5 F 0.0
10:00 230.2 F 0.0
15:00 224.2 F 0.0
20:00 224.9 F 0.0
25:00 229.0 F 0.0
30:00 230.1 F 0.0

Drying Complete

FILLING: 1:60
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

09/20/2022 03:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____