
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002215
10:16 05/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:27
PRESSURIZING 3:32
STERILIZING 4:47
136.2 C 222kPa 4:47
136.3 C 225kPa 5:17
136.2 C 223kPa 5:47
136.4 C 225kPa 6:17
136.4 C 225kPa 6:47
136.2 C 224kPa 7:17
136.2 C 224kPa 7:47
136.4 C 224kPa 8:17
136.4 C 224kPa 8:17
VENTING 8:17
AIR DRYING 9:10
CYCLE COMPLETE 9:12

Digital Signature #
F786885AD1980EF8
