
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000961
11:36 10/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:43
136.5 C 223kPa 3:43
136.2 C 220kPa 4:13
136.3 C 223kPa 4:43
136.4 C 224kPa 5:13
136.4 C 222kPa 5:43
136.4 C 224kPa 6:13
136.4 C 225kPa 6:43
136.4 C 221kPa 7:13
136.4 C 221kPa 7:13
VENTING 7:14
AIR DRYING 8:06
CYCLE COMPLETE 8:14

Digital Signature #
3E1619DE8DF5A93E
