
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002339
7:53 02/09/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:25
STERILIZING 3:16
136.0 C 220kPa 3:16
135.9 C 221kPa 3:46
136.2 C 222kPa 4:16
136.4 C 223kPa 4:46
136.4 C 221kPa 5:16
136.4 C 221kPa 5:46
136.4 C 220kPa 6:16
136.4 C 221kPa 6:46
136.4 C 221kPa 6:46
VENTING 6:46
AIR DRYING 7:37
CYCLE COMPLETE 7:40

Digital Signature #
A879B4333F635FB5
