

-----  
Midmark M11 - v1.0.0  
Total Cycles: 780

Name: RITTER 1  
Sterilizer ID: V2178198  
04/08/2021 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.7 F 0.0  
2:00 100.2 F 0.0  
4:00 110.5 F 0.1  
6:00 139.8 F 0.9  
8:00 174.7 F 4.0  
10:00 200.7 F 7.6  
12:00 219.7 F 11.9  
14:00 222.1 F 6.5  
16:00 235.5 F 11.5  
18:00 233.2 F 8.0  
20:00 242.5 F 11.5  
22:00 243.9 F 12.3  
24:00 256.0 F 18.7  
26:00 267.0 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.9 F 29.2  
2:00 271.6 F 28.3  
3:00 272.1 F 29.1  
4:00 272.0 F 28.8

Min 271.2 F 28.2  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 227.6 F 0.0  
10:00 234.7 F 0.0  
15:00 238.7 F 0.0  
20:00 227.1 F 0.0  
25:00 229.4 F 0.0  
30:00 227.9 F 0.0

Drying Complete

FILLING: 0:59  
HEATING: 27:30  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:05:31  
CYCLE COMPLETE

04/08/2021 01:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_