
Midmark M11 - v1.0.5
Total Cycles: 119

Name: RITTER2
Sterilizer ID: V2313228
06/18/2021 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.8 F 0.0
2:00 114.2 F 0.9
4:00 118.8 F 3.0
6:00 142.7 F 6.0
8:00 171.8 F 9.6
10:00 193.3 F 13.4
12:00 207.8 F 5.7
14:00 219.9 F 9.9
16:00 230.0 F 9.7
18:00 237.6 F 11.8
20:00 238.5 F 10.6
22:00 250.4 F 16.4
24:00 261.0 F 22.5
26:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.0
1:00 272.4 F 30.0
2:00 272.3 F 30.4
3:00 272.2 F 30.5
4:00 271.6 F 29.9

Min 271.6 F 29.8
Max 272.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 216.6 F 0.0
10:00 230.2 F 0.0
15:00 232.1 F 0.0
20:00 225.0 F 0.0
25:00 219.4 F 0.0
30:00 214.1 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 27:30
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:06:06
CYCLE COMPLETE

06/21/2021 06:10 AM
ACCEPTED: X
REJECTED:
Operator: U01