

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4223

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2024 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.6 F 0.0  
2:00 139.4 F 1.7  
4:00 158.4 F 4.5  
6:00 187.5 F 8.0  
8:00 196.8 F 11.8  
10:00 213.5 F 3.4  
12:00 224.8 F 6.9  
14:00 231.4 F 10.7  
16:00 233.3 F 8.1  
18:00 240.5 F 10.9  
20:00 236.3 F 8.3  
22:00 247.6 F 13.6  
24:00 257.3 F 18.8  
26:00 267.1 F 24.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.9  
1:00 272.5 F 28.6  
2:00 271.8 F 28.0  
3:00 271.1 F 28.1  
4:00 272.5 F 28.7

Min 270.9 F 27.7  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.2 F 0.0  
5:00 231.9 F 0.0  
10:00 221.9 F 0.0  
15:00 218.7 F 0.0  
20:00 206.9 F 0.0  
25:00 228.9 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 27:33  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:29  
CYCLE COMPLETE

01/24/2024 12:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_