
Midmark M11 - v1.0.0
Total Cycles: 3600

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.9 F 0.0
2:00 197.0 F 4.0
4:00 216.4 F 9.8
6:00 224.1 F 5.8
8:00 233.6 F 11.0
10:00 233.6 F 8.2
12:00 240.4 F 10.3
14:00 247.4 F 13.8
16:00 259.8 F 20.8
18:00 270.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 29.1
2:00 271.9 F 28.9
3:00 272.3 F 29.0
4:00 271.9 F 28.9

Min 271.6 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 223.8 F 0.0
10:00 226.2 F 0.0
15:00 220.5 F 0.0
20:00 209.3 F 0.0
25:00 207.5 F 0.0
30:00 204.7 F 0.0

Drying Complete

FILLING: 2:28
HEATING: 18:43
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

08/01/2023 12:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____