

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3194

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2023 11:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 77.8 F 0.0  
4:00 76.4 F 0.2  
6:00 97.6 F 1.0  
8:00 132.2 F 3.2  
10:00 163.0 F 6.2  
12:00 176.0 F 9.8  
14:00 194.9 F 13.3  
16:00 213.8 F 5.8  
18:00 222.4 F 9.6  
20:00 236.2 F 12.2  
22:00 234.7 F 9.8  
24:00 238.2 F 9.8  
26:00 244.6 F 13.2  
28:00 255.7 F 19.1  
30:00 265.9 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 273.0 F 29.8  
2:00 271.3 F 29.3  
3:00 272.9 F 29.8  
4:00 271.4 F 29.2

Min 271.2 F 29.1  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 230.1 F 0.0  
10:00 205.4 F 0.0  
15:00 186.8 F 0.0  
20:00 183.1 F 0.0  
25:00 181.6 F 0.0  
30:00 180.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:10:01  
CYCLE COMPLETE

12/07/2023 12:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_