
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000947
16:04 23/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:01
STERILIZING 4:12
136.0 C 220kPa 4:12
136.1 C 230kPa 4:42
136.3 C 232kPa 5:12
136.4 C 233kPa 5:42
136.2 C 231kPa 6:12
136.4 C 233kPa 6:42
136.4 C 232kPa 7:12
136.4 C 231kPa 7:42
136.4 C 231kPa 7:42
VENTING 7:42
AIR DRYING 8:36
CYCLE COMPLETE 20:27

Digital Signature #
A5E8D490D5945236
