

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3212

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2023 03:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.5 F 0.0  
2:00 90.1 F 0.2  
4:00 103.3 F 1.3  
6:00 135.3 F 4.0  
8:00 164.3 F 7.4  
10:00 189.1 F 11.1  
12:00 209.5 F 4.7  
14:00 219.5 F 9.4  
16:00 228.2 F 9.8  
18:00 236.0 F 11.2  
20:00 236.1 F 9.2  
22:00 250.2 F 15.7  
24:00 261.6 F 22.3  
26:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.8  
1:00 272.3 F 30.1  
2:00 272.5 F 30.3  
3:00 272.6 F 30.3  
4:00 272.4 F 30.3

Min 272.0 F 29.7  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
5:00 210.6 F 0.0  
10:00 208.6 F 0.0  
15:00 201.9 F 0.0  
20:00 200.1 F 0.0  
25:00 201.6 F 0.0  
30:00 212.6 F 0.0

Drying Complete

FILLING: 1:02  
HEATING: 27:13  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:05:28  
CYCLE COMPLETE

12/13/2023 04:31 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_