

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001745  
8:31 14/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:48  
STERILIZING 3:55  
136.0 C 220kPa 3:55  
135.9 C 220kPa 4:25  
136.2 C 224kPa 4:55  
135.9 C 220kPa 5:25  
136.2 C 222kPa 5:55  
136.2 C 222kPa 6:25  
CYCLE ABORTED 6:28  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 6:28  
-----

Digital Signature #  
ED1D2FC3CF5A6370  
-----