
Midmark M11 - v1.0.5
Total Cycles: 2366

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 11:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.1 F 0.0
2:00 198.9 F 3.8
4:00 208.0 F 9.4
6:00 223.0 F 7.8
8:00 234.9 F 11.2
10:00 235.0 F 9.2
12:00 237.9 F 9.5
14:00 249.3 F 15.3
16:00 261.6 F 22.3
18:00 272.8 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.5
2:00 271.3 F 28.7
3:00 272.5 F 29.6
4:00 272.0 F 29.1

Min 271.3 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 210.2 F 0.0
10:00 227.5 F 0.0
15:00 227.9 F 0.0
20:00 227.3 F 0.0
25:00 228.0 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

05/15/2023 12:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____