

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007879  
8:06 25/09/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:27  
STERILIZING 3:34  
136.4 C 223kPa 3:34  
136.2 C 221kPa 4:04  
136.5 C 223kPa 4:34  
136.2 C 217kPa 5:04  
136.2 C 217kPa 5:34  
136.4 C 221kPa 6:04  
136.4 C 223kPa 6:34  
136.4 C 223kPa 7:04  
136.4 C 223kPa 7:04  
VENTING 7:04  
AIR DRYING 7:57  
CYCLE COMPLETE 20:18

-----

Digital Signature #  
DD45B7D95C19A013  
-----