

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1784

Name: RITTER 1  
Sterilizer ID: V2178198  
04/01/2022 12:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
2:00 173.6 F 3.7  
4:00 193.7 F 8.0  
6:00 217.7 F 10.8  
8:00 230.3 F 8.5  
10:00 239.0 F 11.1  
12:00 240.7 F 11.6  
14:00 239.5 F 10.2  
16:00 253.3 F 17.1  
18:00 265.2 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.5  
1:00 272.2 F 28.7  
2:00 272.6 F 29.3  
3:00 271.1 F 28.6  
4:00 272.4 F 28.9  
  
Min 271.0 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 222.7 F 0.0  
10:00 204.4 F 0.0  
15:00 193.5 F 0.0  
20:00 189.0 F 0.0  
25:00 193.4 F 0.0  
30:00 233.8 F 0.0  
Drying Complete

FILLING: 2:06  
HEATING: 19:41  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:56  
CYCLE COMPLETE

04/01/2022 01:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_