
Midmark M11 - v1.0.5
Total Cycles: 4481

Name: RITTER2
Sterilizer ID: V2313228
04/03/2025 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 79.1 F 0.1
4:00 111.5 F 0.9
6:00 108.9 F 2.9
8:00 149.0 F 5.8
10:00 170.4 F 9.3
12:00 189.7 F 13.4
14:00 218.9 F 4.9
16:00 227.5 F 8.3
18:00 236.7 F 12.2
20:00 230.5 F 7.9
22:00 242.1 F 12.9
24:00 239.7 F 10.9
26:00 251.7 F 16.8
28:00 262.3 F 23.0
30:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.1 F 29.2
2:00 272.6 F 29.6
3:00 271.9 F 29.4
4:00 272.0 F 29.0

Min 271.5 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 179.0 F 0.0
10:00 192.4 F 0.0
15:00 197.0 F 0.0
20:00 199.4 F 0.0
25:00 201.4 F 0.0
30:00 202.8 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

04/03/2025 10:02 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____