

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2741

Name: RITTER 1  
Sterilizer ID: V2178198  
01/05/2023 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.8 F 0.0  
2:00 201.9 F 0.0  
4:00 213.4 F 4.2  
6:00 227.5 F 9.1  
8:00 235.1 F 9.6  
10:00 236.4 F 10.5  
12:00 237.9 F 7.9  
14:00 243.1 F 11.4  
16:00 245.9 F 13.1  
18:00 258.4 F 19.8  
20:00 269.5 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.8  
2:00 271.7 F 28.6  
3:00 272.2 F 28.6  
4:00 271.7 F 28.5  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 225.2 F 0.0  
10:00 238.9 F 0.0  
15:00 232.3 F 0.0  
20:00 221.9 F 0.0  
25:00 218.5 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 21:00  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:54  
CYCLE COMPLETE

01/05/2023 12:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_