

-----  
Midmark M11 - v1.0.0  
Total Cycles: 897

Name: RITTER 1  
Sterilizer ID: V2178198  
06/15/2021 08:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.1 F 0.0  
2:00 121.9 F 0.0  
4:00 136.5 F 0.1  
6:00 171.8 F 1.5  
8:00 195.8 F 5.2  
10:00 212.2 F 9.0  
12:00 224.1 F 13.2  
14:00 228.8 F 7.9  
16:00 240.5 F 12.9  
18:00 238.7 F 10.3  
20:00 235.8 F 8.2  
22:00 250.0 F 15.2  
24:00 262.0 F 22.0  
26:00 272.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.2  
1:00 272.6 F 29.0  
2:00 271.8 F 28.2  
3:00 272.6 F 29.0  
4:00 272.1 F 28.4

Min 271.5 F 28.1  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.6 F 0.0  
5:00 212.1 F 0.0  
10:00 220.7 F 0.0  
15:00 228.9 F 0.0  
20:00 230.6 F 0.0  
25:00 225.6 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 26:14  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:04:26  
CYCLE COMPLETE

06/15/2021 09:34 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_