
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 004199
8:43 16/08/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 4:42
STERILIZING 5:30
136.4 C 220kPa 5:30
CYCLE ABORTED 5:43
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:43

Digital Signature #
5BC67632914BA186
