
Midmark M11 - v1.0.0
Total Cycles: 3282

Name: RITTER 1
Sterilizer ID: V2178198
05/25/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.1 F 0.0
2:00 193.1 F 0.0
4:00 187.3 F 2.4
6:00 207.5 F 7.6
8:00 221.6 F 13.0
10:00 231.2 F 8.5
12:00 240.0 F 11.3
14:00 243.4 F 12.3
16:00 242.3 F 11.2
18:00 255.8 F 18.2
20:00 267.7 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 29.0
2:00 272.3 F 28.5
3:00 271.5 F 28.0
4:00 272.0 F 28.6

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 237.2 F 0.0
10:00 229.8 F 0.0
15:00 231.5 F 0.0
20:00 227.2 F 0.0
25:00 227.6 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

05/25/2023 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____