

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4572

Name: RITTER 1  
Sterilizer ID: V2178198  
05/22/2024 11:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.7 F 0.0  
2:00 103.0 F 0.8  
4:00 127.9 F 2.8  
6:00 152.5 F 5.4  
8:00 174.3 F 8.0  
10:00 194.4 F 11.4  
12:00 210.2 F 3.2  
14:00 216.5 F 6.1  
16:00 223.6 F 9.8  
18:00 235.8 F 11.4  
20:00 235.2 F 9.9  
22:00 239.8 F 9.4  
24:00 245.1 F 13.2  
26:00 255.9 F 18.8  
28:00 265.8 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.1 F 29.4  
2:00 272.5 F 29.3  
3:00 271.7 F 28.8  
4:00 271.1 F 28.7

Min 271.1 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 202.4 F 0.0  
10:00 200.0 F 0.0  
15:00 202.2 F 0.0  
20:00 218.8 F 0.0  
25:00 219.9 F 0.0  
30:00 221.6 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 29:56  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:08:39  
CYCLE COMPLETE

05/22/2024 01:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_