
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006927
8:59 12/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:54
STERILIZING 3:04
136.5 C 223kPa 3:04
136.4 C 224kPa 3:34
136.1 C 221kPa 4:04
136.4 C 222kPa 4:34
136.1 C 222kPa 5:04
136.3 C 224kPa 5:34
136.4 C 223kPa 6:04
136.4 C 224kPa 6:34
136.4 C 224kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 17:22

Digital Signature #
DD1FDF451A36A159
