
Midmark M11 - v1.0.5
Total Cycles: 984

Name: RITTER2
Sterilizer ID: V2313228
04/13/2022 12:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.4 F 0.0
2:00 156.4 F 0.3
4:00 183.3 F 3.2
6:00 207.0 F 8.7
8:00 219.4 F 8.5
10:00 233.5 F 10.9
12:00 234.9 F 9.0
14:00 238.4 F 9.6
16:00 249.9 F 15.7
18:00 263.1 F 23.3
20:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.6 F 29.3
2:00 272.4 F 29.7
3:00 272.4 F 29.1
4:00 272.5 F 29.8

Min 271.8 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 207.3 F 0.0
10:00 215.3 F 0.0
15:00 229.0 F 0.0
20:00 228.7 F 0.0
25:00 224.1 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:02
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 00:58:07
CYCLE COMPLETE

04/13/2022 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____