
Midmark M11 - v1.0.5
Total Cycles: 868

Name: RITTER2
Sterilizer ID: V2313228
03/10/2022 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 73.7 F 0.0
4:00 76.6 F 0.0
6:00 104.0 F 0.1
8:00 137.4 F 1.6
10:00 166.2 F 4.4
12:00 184.7 F 7.5
14:00 201.7 F 11.3
16:00 216.0 F 5.2
18:00 231.1 F 9.9
20:00 234.8 F 9.5
22:00 240.4 F 11.6
24:00 237.9 F 9.8
26:00 251.0 F 16.1
28:00 261.9 F 22.5
30:00 272.2 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.7 F 29.7
2:00 272.0 F 29.0
3:00 271.5 F 29.0
4:00 272.7 F 29.7

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 198.4 F 0.0
10:00 197.9 F 0.0
15:00 175.3 F 0.0
20:00 162.7 F 0.0
25:00 165.0 F 0.0
30:00 156.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:41
CYCLE COMPLETE

03/10/2022 08:51 AM
ACCEPTED: X
REJECTED:
Operator: U02