

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3215

Name: RITTER2  
Sterilizer ID: V2313228  
12/14/2023 11:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.1 F 0.0  
2:00 152.1 F 1.4  
4:00 167.2 F 4.4  
6:00 190.3 F 9.4  
8:00 221.7 F 7.4  
10:00 229.1 F 9.3  
12:00 234.9 F 9.3  
14:00 241.5 F 11.8  
16:00 239.2 F 10.2  
18:00 252.8 F 16.9  
20:00 264.6 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.7  
1:00 273.2 F 29.7  
2:00 271.9 F 29.5  
3:00 272.4 F 29.2  
4:00 273.0 F 29.9

Min 271.0 F 28.6  
Max 273.3 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.1 F 0.0  
5:00 234.1 F 0.0  
10:00 210.7 F 0.0  
15:00 189.1 F 0.0  
20:00 200.4 F 0.0  
25:00 206.4 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:55  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:13  
CYCLE COMPLETE

12/14/2023 12:53 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_