
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001595
15:17 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:08
STERILIZING 3:15
136.4 C 226kPa 3:15
136.2 C 225kPa 3:45
136.2 C 224kPa 4:15
136.3 C 225kPa 4:45
136.0 C 223kPa 5:15
136.4 C 226kPa 5:45
136.4 C 225kPa 6:15
136.4 C 225kPa 6:45
136.4 C 225kPa 6:45
VENTING 6:45
AIR DRYING 7:37
CYCLE COMPLETE 7:42

Digital Signature #
518DDAA0F2FFAFA1
