
Midmark M11 - v1.0.0
Total Cycles: 2917

Name: RITTER 1
Sterilizer ID: V2178198
02/16/2023 05:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 113.4 F 1.2
4:00 131.9 F 3.7
6:00 159.0 F 6.9
8:00 198.9 F 10.9
10:00 206.9 F 3.3
12:00 217.1 F 7.7
14:00 228.7 F 11.9
16:00 229.1 F 7.3
18:00 240.3 F 12.5
20:00 238.8 F 10.1
22:00 251.1 F 16.1
24:00 262.2 F 22.4
26:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.8 F 29.6
2:00 271.6 F 28.8
3:00 272.5 F 29.6
4:00 272.3 F 29.2

Min 271.4 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 210.3 F 0.0
10:00 221.7 F 0.0
15:00 230.1 F 0.0
20:00 235.3 F 0.0
25:00 230.5 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:34
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

02/16/2023 06:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____