
Midmark M11 - v1.0.0
Total Cycles: 3379

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2023 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 193.4 F 0.5
4:00 207.8 F 5.7
6:00 222.9 F 11.1
8:00 228.3 F 7.0
10:00 239.0 F 12.6
12:00 239.3 F 10.6
14:00 237.0 F 9.2
16:00 252.1 F 16.4
18:00 264.1 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.4 F 28.9
2:00 271.5 F 28.4
3:00 272.5 F 29.0
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 205.9 F 0.0
10:00 225.8 F 0.0
15:00 230.3 F 0.0
20:00 230.2 F 0.0
25:00 225.8 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

06/13/2023 10:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____