
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000737
8:27 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:14
STERILIZING 3:17
136.0 C 227kPa 3:17
136.4 C 233kPa 3:47
136.4 C 230kPa 4:17
136.1 C 229kPa 4:47
136.4 C 233kPa 5:17
136.3 C 233kPa 5:47
136.4 C 231kPa 6:17
136.4 C 233kPa 6:47
136.4 C 233kPa 6:47
VENTING 6:47
AIR DRYING 7:40
CYCLE COMPLETE 8:35

Digital Signature #
C8039A38D90247D2
