
Midmark M11 - v1.0.0
Total Cycles: 5020

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2024 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.5 F 0.0
2:00 110.6 F 0.0
4:00 122.3 F 0.2
6:00 150.3 F 0.8
8:00 178.3 F 1.9
10:00 197.1 F 5.3
12:00 211.9 F 8.7
14:00 223.6 F 12.5
16:00 224.8 F 6.9
18:00 236.3 F 11.6
20:00 234.6 F 8.5
22:00 242.5 F 11.5
24:00 244.9 F 12.6
26:00 256.1 F 18.5
28:00 266.6 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.3 F 28.9
2:00 271.9 F 28.4
3:00 271.1 F 28.1
4:00 272.1 F 28.9

Min 271.0 F 28.0
Max 272.3 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 193.0 F 0.0
10:00 186.4 F 0.0
15:00 189.3 F 0.0
20:00 193.5 F 0.0
25:00 207.1 F 0.0
30:00 213.6 F 0.0
Drying Complete

FILLING: 0:37
HEATING: 29:44
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:08:11
CYCLE COMPLETE

10/24/2024 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____