
Midmark M11 - v1.0.5
Total Cycles: 1633

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.3 F 0.0
2:00 185.0 F 4.4
4:00 207.0 F 10.5
6:00 229.2 F 7.5
8:00 240.1 F 10.9
10:00 243.8 F 13.1
12:00 244.1 F 12.7
14:00 258.3 F 20.4
16:00 270.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.8 F 30.1
2:00 272.6 F 30.1
3:00 272.6 F 30.2
4:00 271.9 F 29.9

Min 270.7 F 28.9
Max 273.3 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 238.4 F 0.0
10:00 229.0 F 0.0
15:00 233.9 F 0.0
20:00 230.5 F 0.0
25:00 236.3 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 3:37
HEATING: 16:38
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 00:57:16
CYCLE COMPLETE

10/20/2022 02:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____