
Midmark M11 - v1.0.0
Total Cycles: 2820

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.3 F 0.0
2:00 149.0 F 0.0
4:00 171.3 F 3.1
6:00 192.3 F 7.5
8:00 214.4 F 12.6
10:00 223.3 F 7.2
12:00 237.2 F 12.4
14:00 236.1 F 9.3
16:00 239.2 F 9.6
18:00 247.4 F 13.8
20:00 259.2 F 20.3
22:00 269.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.6 F 28.9
2:00 271.7 F 28.2
3:00 272.5 F 29.0
4:00 272.2 F 28.6

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 217.4 F 0.0
10:00 229.5 F 0.0
15:00 230.7 F 0.0
20:00 231.4 F 0.0
25:00 219.9 F 0.0
30:00 217.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 22:59
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

01/25/2023 02:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____