
Midmark M11 - v1.0.5
Total Cycles: 1597

Name: RITTER2
Sterilizer ID: V2313228
10/14/2022 09:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 167.8 F 2.5
4:00 184.1 F 5.6
6:00 201.6 F 9.9
8:00 216.7 F 6.6
10:00 221.0 F 8.3
12:00 234.0 F 13.3
14:00 236.1 F 10.5
16:00 234.8 F 8.3
18:00 247.9 F 15.1
20:00 259.8 F 21.6
22:00 270.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.6
1:00 272.5 F 29.9
2:00 271.9 F 29.9
3:00 272.3 F 29.8
4:00 272.0 F 30.1

Min 271.6 F 29.4
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 220.9 F 0.0
10:00 234.6 F 0.0
15:00 231.5 F 0.0
20:00 228.3 F 0.0
25:00 228.2 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:52
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

10/14/2022 10:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____