
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 003658
8:49 12/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:45
STERILIZING 3:55
136.0 C 220kPa 3:55
136.1 C 222kPa 4:25
136.4 C 223kPa 4:55
136.1 C 223kPa 5:25
136.4 C 224kPa 5:55
136.3 C 223kPa 6:25
136.3 C 222kPa 6:55
136.4 C 223kPa 7:25
136.4 C 223kPa 7:25
VENTING 7:26
AIR DRYING 8:15
CYCLE COMPLETE 8:21

Digital Signature #
909A3B220156220E
