
Midmark M11 - v1.0.5
Total Cycles: 1679

Name: RITTER2
Sterilizer ID: V2313228
11/03/2022 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.2 F 0.0
2:00 185.2 F 3.6
4:00 203.6 F 9.1
6:00 222.4 F 7.9
8:00 233.7 F 11.0
10:00 234.8 F 9.1
12:00 238.2 F 9.5
14:00 249.7 F 15.4
16:00 262.0 F 22.4
18:00 273.0 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.3
2:00 271.3 F 28.9
3:00 272.6 F 29.4
4:00 271.4 F 28.7

Min 271.3 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 228.1 F 0.0
10:00 243.3 F 0.0
15:00 237.6 F 0.0
20:00 225.7 F 0.0
25:00 220.4 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:53
CYCLE COMPLETE

11/03/2022 12:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____