

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001051  
12:08 29/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:32  
STERILIZING 3:29  
136.4 C 222kPa 3:29  
136.3 C 222kPa 3:59  
136.2 C 224kPa 4:29  
136.3 C 224kPa 4:59  
136.4 C 224kPa 5:29  
136.3 C 223kPa 5:59  
136.4 C 222kPa 6:29  
136.4 C 223kPa 6:59  
136.4 C 223kPa 6:59  
VENTING 6:59  
AIR DRYING 7:53  
CYCLE COMPLETE 8:23

-----  
Digital Signature #  
576D92E9314EACFC  
-----